

## Procedure


Extraction

Hydrolysis


Purification

Measurement


### YMC-Triart C18 (1.9 $\mu\text{m}$ , 12 nm)




### YMC-Triart C18 ExRS (1.9 $\mu\text{m}$ , 8 nm)




### YMC-Triart C8 (1.9 $\mu\text{m}$ , 12 nm)




### YMC-Triart Phenyl (1.9 $\mu\text{m}$ , 12 nm)



### YMC-Triart PFP (1.9 $\mu\text{m}$ , 12 nm)



### Okadaic acid



	R <sub>1</sub>	R <sub>2</sub>
1. Okadaic acid (OA)	CH <sub>3</sub>	H
2. Dinophysistoxin-1 (DTX1)	CH <sub>3</sub>	CH <sub>3</sub>
3. Dinophysistoxin-2 (DTX2)	H	CH <sub>3</sub>

Courtesy of Japan Frozen Foods Inspection Corporation

Column : 75 X 2.1 mm I.D.

Eluent : A) water containing 2 mM HCOONH<sub>4</sub> and 50 mM HCOOH  
 B) acetonitrile/water (95/5) containing 2 mM HCOONH<sub>4</sub> and 50 mM HCOOH  
 40%B (0-2.5 min), 40%-100%B (2.5-7.5 min), 100%B (7.5-12.5 min)

Temperature : 40°C

Flow rate : 0.2 mL/min

Detection : AB SCIEX QTRAP® 4500, ESI, Negative, MRM

Okadaic acid (*m/z* 803>255)

Dinophysistoxin-1 (*m/z* 817>255)

Dinophysistoxin-2 (*m/z* 803>255)

Sample : Matrix matched standards

Injection : 5  $\mu\text{L}$  (5 ppb)