



Courtesy of Dr. Toyooki Aoki, University of Osaka Pref.

Column	: YMC-Pack ODS-A (5 $\mu$ m, 12nm) 150 $\times$ 6.0 mm I.D.
Eluent	: methanol
Flow rate	: 1.0 mL/min
Temperature	: ambient
Detection	: VIS at 450 nm, 1 AUFS
Injection	: 100 $\mu$ L
Sample	: extract of Spinacia with acetone/water (90/10)